

Synthetic Blocking Buffer, ELISA

BUFFERS & STABILIZERS

Microwell Blocker Cat. No. 4520

Lot No.: <u>xxxxxxXxx</u>

Date of manufacture: <u>mm-yyyy</u>

Expiry date: mm-yyyy (When stored at 2-8°C in an unopened bottle)

Analysis: The non-specific binding (NSB) of the blocking buffer is analyzed in a 1-layer ELISA. It is

furthermore analyzed for consistent pH and appearance. Only the procedure for the

ELISA analysis is described in the CoA.

NSB test: All wells are blocked by addition of 300 μ L Synthetic Blocking Buffer, ELISA. A solution of Rabbit-anti-human IgG-HRP is added the wells. TMB PLUS2 is added and incubated for 13 minutes. The reaction is stopped by addition of 0.2 M sulfuric acid.

The absorbance is read at 450/620 nm.

Specifications: NSB: <0.150 Abs.

Appearance: Clear solution with no distinct color

pH: 7.1 - 7.3

Results: NSB: <u>x.xxx Abs.</u>

Appearance: <u>Passed</u> pH: <u>x.xx</u>

Storage: Short-term storage: Ambient; protected from light.

Long-term storage: 2-8 °C; protected from light.

It is hereby confirmed that the product was manufactured according to our standard operating procedures and that the product complies with all specifications.

Performed by: Date: <u>dd-MMM-yyyy</u> Signature: <u>XX</u>

Approved by: Date: dd-MMM-yyyy Signature: XX

Our Immunoassay Solutions are developed in order to create a healthy and safe work environment and preserving natural resources, while helping our customers fulfill significant regulatory requirements. **We call it ECO-TEK!**

