## kementec

## TMB EASE

		CHROMOGENIC SUBSTRATES		
ELISA HRP Substrate Cat. No. 5320				
Lot No.:	<u>xxxxxXxx</u>			
Date of manufacture:	<u>mm-yyyy</u>			
Expiry date:	<u>mm-yyyy</u>	(When stored at 2-8°C in an unopened bottle)		
Analysis:	<ul> <li>The substrate is analyzed for consistent activity, background, appearance, and pH. Only the procedure for activity and background is described in the CoA.</li> <li>Background: 100 μL substrate is pipetted into four wells of a standard microtiter plate, prefilled with 100 μL of 0.2 M sulphuric acid. The background absorbance is read at 450/620 nm.</li> <li>Activity: The activity is determined in a 2-layer ELISA. First layer is a human lgG antibody dilution. Second layer is Rabbit-anti-human IgG-HRP. 100 μL substrate is added to each well and incubated for 13 minutes. The reaction is stopped by addition of 100 μL 0.2 M sulphuric acid. The absorbance is read at 450/620 nm.</li> </ul>			
Specifications:	Activity: Background: Appearance: pH:	90 – 110% < 0.025 Abs. Clear solution with no distinct color 3.50 – 3.90		
Results:	Activity: Background: Appearance: pH:	<u>xxx%</u> x.xxx Abs. Passed x.xx		
Storage:		prage: Ambient; protected from light. rage: 2-8 °C; protected from light.		

It is hereby confirmed that the product was manufactured according to our standard operating procedures and that the product complies with all specifications.

Performed by:	Date: dd-MMM-yyyy	Signature: XX
Approved by:	Date: <u>dd-MMM-yyyy</u>	Signature: XX

Our Immunoassay Solutions are developed in order to create a healthy and safe work environment and preserving natural resources, while helping our customers fulfill significant regulatory requirements. We call it ECO-TEK!



Kuldyssen 10 DK-2630 Taastrup Denmark VAT: DK 25 80 67 35

Tel: +45 39 27 17 71 Fax: +45 39 20 01 78