kementec

TMB ONE

		CHROMOGENIC SUBSTRATES			
ELISA HRP Substrate					
Cat. No. 4380					
Lot No.:	xxxxXxx				
Date of manufacture:	<u>mm-yyyy</u>				
Expiry date:	<u>mm-yyyy</u>		(When stored at 2-8°C in an ι	inopened bottle)	
Analysis:	The substrate is analyzed for consistent activity, background, appearance, and pH. Only the procedure for activity and background is described in the CoA.				
	Background: 100 μ L substrate is pipetted into four wells of a standard microtiter plate, prefilled with 100 μ L of 0.2 M sulphuric acid. The background absorbance is read at 450/620 nm.				
	Activity: The activity is determined in a 2-layer ELISA. First layer is a human IgG antibody dilution. Second layer is Rabbit-anti-human IgG-HRP. 100 μ L substrate is added to each well and incubated for 13 minutes. The reaction is stopped by addition of 100 μ L 0.2 M sulphuric acid. The absorbance is read at 450/620 nm and converted to activity using a standard curve.				
Specifications:	Activity: Background: Appearance: pH:	95 – 105% < 0.025 Abs. Clear solution with no o 3.80 – 4.20	listinct color		
Results:	Activity: Background: Appearance: pH:	<u>xxx%</u> x.xxx Abs. Passed x.xx			
Storage:	Short-term storage: Ambient; protected from light. Long-term storage: 2-8 °C; protected from light.				
It is hereby confirmed that the product was manufactured according to our standard operating procedures and that the product complies with all specifications.					
Performed by:	Date: dd-MMM-yyyy		Signature: XX		
Approved by:	Date: dd-MMM-yyyy		Signature: <u>XX</u>		
Our Immunoassay Solutions are developed in order to create a healthy and safe work environment and preserving natural resources, while helping our customers fulfill significant regulatory requirements. We call it ECO-TEK!					

Kuldyssen 10 DK-2630 Taastrup | Denmark | VAT: DK 25 80 67 35 Tel: +45 39 27 17 71 Fax: +45 39 20 01 78 | support@kementec.com | www.kementec.com